

Area: Logistics	SME:	DATE:
------------------------	------	-------

Responsibility (L7): Ensure radiacmeter calibration and/or certification is current.

Steps / Procedures	Local POC	Remarks/References
<p>*Maintain a chart that lists all the NBC Defense Equipment requiring calibration or wipe testing .</p> <p>As a minimum the chart should include:</p> <ol style="list-style-type: none"> Name of equipment Date calibrated Date due calibration Wipe test date Date wipe test is due <p>*Ensure that this chart is always updated as equipment is turned in for service or replacement.</p>	<p>Co NBC Off; BN NBC Off/NCO; Supply NCO</p>	<p>Maintenance Update; TMs; TMDE SB 43-180</p>
<p>*PMCS and prepare DA Form 2404.</p> <p>*Prepare DA Form 2407 Work Request listing required information on equipment being turned-in for calibration</p> <p>*Turn items in IAW Unit SOP</p> <p>*Receive equipment back from Maintenance unit</p>	<p>Co NBC Off; BN NBC Off/NCO; Supply NCO</p>	<p>Maintenance Update; TMs; TMDE SB 43-180</p>
<p>*Track paperwork, keep your copy of DA Form 2407 Work Request to prove item has been calibrated.</p>	<p>Co NBC Off; BN NBC Off/NCO; Supply NCO</p>	<p>Maintenance Update; TMs; TMDE SB 43-180</p>
<p>*Annotate calibration data on DD Form 314 Scheduled Maintenance.</p>	<p>Co NBC Off; BN NBC Off/NCO; Supply NCO</p>	<p>Maintenance Update; TMs; TMDE SB 43-180</p>