

TEST, MEASUREMENT, AND DIAGNOSTIC EQUIPMENT (TMDE)

Proponent For Inspection: DOL

Point of Contact: _____

Unit Inspected: _____

Date of Inspection: _____

Unit Representative: _____

Unit Phone No.: _____

Inspector's Name: _____

Inspector's Phone No.: _____

Unit Overall Rating: T P U

- REFERENCES: a. AR 750-1 W/Ch 1, 1 Jul 96 b. AR 750-43, 28 Nov 02 c. TB 43-180, 15 Sep 02
 d. TB 750-25 W/CH 1, 12 Nov 99 e. DA Pam 738-750, 1 Aug 94 f. DA Pam 750-35, 1 Aug 94
 g. TSC-FLW SOP #738-1, 22 Jul 02 h. Current applicable TMs for the equipment being inspected

STANDARDS: "T"= 90% success rate of evaluated tasks with no failed critical tasks. "P"= 70% success rate of evaluated tasks with no failed critical tasks. "U"= less than 70% success rate of evaluated tasks or one failed critical task

INSPECTION CRITERIA:	LEVEL	GO	NO GO	REMARKS
1. Has the commander appointed in writing a designated representative as TMDE Support Coordinator to serve as the central point of contact between the unit and the TMDE Support Activity (TSA)? (AR 750-43, Para 6-4; TB 750-25, Para 3-8b, Para 3-10a a (1))	ALL			
2. As a minimum, does the TMDE coordinator have the following publications: AR 750-1, AR 750-43, TB 750-25, TB 43-180, DA Pam 738-750, and the TSA External SOP? (TB 750-25, Para 3-10a(3))	CO/BN			
3. CRITICAL: Has the organization met the minimum DA goal of having 95% or better of the TMDE inventory calibrated and in repaired condition and available to the user, with no more than a 3% delinquency rate? (AR 750-43, Para 6-1a & b)	ALL			
4. Does the TMDE coordinator have a TMDE Master List or Instrument Master Record File (IMRF) provided by the TSA quarterly, identifying all instruments belonging to the user? (AR 750-43, Para 6-12, Para 6-15 e (3); TB 750-25, Para 3-2b(2) & c, Para 3-10a(4))	CO/BN			
5. Does the TMDE coordinator have a Delinquent Items List, provided by the TSA monthly, identifying TMDE not presented for calibration within the scheduled timeframe? Is this list reviewed and is corrective action taken? (AR 750-43, Appendix B Para B-4h; TB 750-25, Para 3-2b(2), Para 3-10a(6), Para 3-10b(7))	CO/BN			
6. Does the TMDE coordinator have a Projected Items List (PIL), provided by the TSA at least monthly (minimum 30-day notice), identifying each instrument in the IRMF due for C&RS within that month? Is this list reviewed for accuracy? (TB 750-25, Para 3-2b, Para 3-10a(5), Para 3-10b(6))	CO/BN			
7. Does TMDE in use have a current DA Label 80 or DA Label 163 affixed and properly annotated? (AR 750-43, Para 6-14; TB 750-25, Para 3-10b(12))	CO/BN			
8. Has an operator/unit level maintenance program been established for TMDE (including CNR labeled TMDE) IAW equipment maintenance manuals? (AR 750-1, Para 5-51a; TB 750-25, Para 3-10b(10) & (11))	CO/BN			
9. Is PMCS being performed and documented, and are records filed or kept with TMDE as appropriate? (AR 750-1, Para 3-9e)	CO/BN			
10. Has an SOP been developed and implemented for the identification and control of TMDE requiring calibration and repair support? (TB 750-25, Para 3-10a(2))	CO/BN			
REMARKS AND FINAL COMMENTS:				